



Randport Industrial  
Unit 1, 1 Suzuka Road  
Gosforth Park  
Germiston  
1401

P.O. Box 1382  
Alberton  
1450

T: +27 11 538 0600  
F: +27 86 435 4399  
Reg. No:  
1992/005895/07  
VAT No: 4080138532

CooperVision South Africa  
09 February 2022

**COMPANY-INITIATED FIELD SAFETY NOTICE**  
**Hy-Care® Multipurpose Contact Lens Care Solution - 60ml & 360ml**

Dear Practitioner,

CooperVision is voluntarily recalling its Hy-Care® contact lens solution from the market, effective immediately. Our quality testing has determined that the product may not offer appropriate levels of disinfection in some circumstances against high levels of a particular yeast organism. Although no complaints have been received and there have been no confirmed incidents attributed to inadequate disinfection, we are acting out of an abundance of caution.

Our records indicate you have received affected product in one or more of the above bottle configurations. We would appreciate your assistance in the following actions:

- Please immediately examine your inventory / product shelf;
- Stop any further distribution of the affected product;
- Record, where practical, the LOT numbers and quantity of each bottle type from your inventory;
- Contact CooperVision Customer Services on (011) 538 0600 to understand how to obtain a credit against affected product;
- Dispose of the solution through waste-water systems and recycle the primary packaging; if you do not wish to dispose of the product in this way, you may return the product to CooperVision through your normal means.

As part of this action, we will be asking consumers to cease use and dispose of any affected product they have. The process will be as follows and we are currently preparing the mechanisms to undertake these actions – we will be in contact with you in due course to provide further information:

- Notification to consumers (email/letter template) – linking to a Recall website;
- Post for your website – linking to the Recall website;
- Recall website – users will be advised to dispose of the solution through their household waste-water system and recycle the primary packaging. There will be instructions on how to obtain reimbursement for the product.

While we prepare the above, we ask that you post the enclosed/attached consumer notification on your website.

We appreciate your assistance and apologize for any inconvenience to you or your customers. Should you wish to discuss this action, please contact your business development manager or our customer services team at: (011) 538 0600

Yours sincerely,

Anton du Preez  
Country Manager

Enclosure/attachment:

Product Recall Notification Template (Customer to Consumer)

Frequently Asked Questions document